



BILEVEL CONTROL

Bilevel Protein control
Immuno Turbidimetry

IVD
Store at 2 – 8 °C

PRODUCT CHARACTERISTICS

This protein control is a human serum, with added serum proteins, constituents of animal origin, stabilizers and preservatives.

PRECAUTIONS

Components from human origin have been tested and found negative for the presence for HbsAg, HCV and antibody to HIV (1/2). However handle cautiously as potentially infectious.

STORAGE AND STABILITY

This control is stable until the expiration date on the label when stored unopened at 2-8 °C.

PROCEDURE

This product should be treated the same as a patient specimen and run in accordance with the instructions accompanying the kit. Before sampling, allow the control to reach room temperature (18 – 25 °C). After each use, promptly replace the stopper and return to 2 – 8 °C storage.

LIMITATIONS

1. This product must not be used past the expiration date
2. If there is evidence of microbial contamination in the product, discard the vial.
3. Complement C4 levels may gradually decrease during the product shelf life.

COMPONENT	Mean	Lev 1	Mean	Lev 2	UNITS
α1-antitrypsin					mg/dL
α1-acid glycoprotein					mg/dL
α2-macroglobulin					mg/dL
Antistreptolysin O					UI/mL
Antithrombin III					mg/dL
β2 Microglobulin					mg/L
C3					mg/dL
C4					mg/dL
C Reactive Protein					mg/L
Haptoglobin					mg/dL
IgA					mg/dL
IgG					mg/dL
IgM					mg/dL
Prealbumin					mg/dL
Transferrin					mg/dL

Ref: ACC16-057 Cont: 2x3 mL Lot: 0708005
Expiry date: 31/07/2010



VITAL DIAGNOSTICS S.r.l. via Balzella 41/G/4, 47100 Forlì, ITALY
Tel: +39 0543 721220 Fax: +39 0543 796001

TF_CTR_BILEVEL_0708005(00).doc



BILEVEL CONTROL

Bilevel Protein control
Immuno Turbidimetry

IVD
Store at 2 – 8 °C

PRODUCT CHARACTERISTICS

This protein control is a human serum, with added serum proteins, constituents of animal origin, stabilizers and preservatives.

PRECAUTIONS

Components from human origin have been tested and found negative for the presence for HbsAg, HCV and antibody to HIV (1/2). However handle cautiously as potentially infectious.

STORAGE AND STABILITY

This control is stable until the expiration date on the label when stored unopened at 2-8 °C.

PROCEDURE

This product should be treated the same as a patient specimen and run in accordance with the instructions accompanying the kit. Before sampling, allow the control to reach room temperature (18 – 25 °C). After each use, promptly replace the stopper and return to 2 – 8 °C storage.

LIMITATIONS

1. This product must not be used past the expiration date
2. If there is evidence of microbial contamination in the product, discard the vial.
3. Complement C4 levels may gradually decrease during the product shelf life.

COMPONENT	Mean	Lev 1	Mean	Lev 2	UNITS
α1-antitrypsin					mg/dL
α1-acid glycoprotein					mg/dL
α2-macroglobulin					mg/dL
Antistreptolysin O					UI/mL
Antithrombin III					mg/dL
β2 Microglobulin					mg/L
C3					mg/dL
C4					mg/dL
C Reactive Protein					mg/L
Haptoglobin					mg/dL
IgA					mg/dL
IgG					mg/dL
IgM					mg/dL
Prealbumin					mg/dL
Transferrin					mg/dL

Ref: ACC16-057 Cont: 2x1 mL Lot: 0708005
Expiry date: 31/07/2010



VITAL DIAGNOSTICS S.r.l. via Balzella 41/G/4, 47100 Forlì, ITALY
Tel: +39 0543 721220 Fax: +39 0543 796001

TF_CTR_BILEVEL_0708005(00).doc